

Declaration of Conformity

We declare, under our sole responsibility, that the product listed below fulfills the requirements specified in Regulation (EU) 2017/746 on in-vitro diagnostic medical devices.

Manufacturer's Name and Business Address:	Leica Biosystems Melbourne Pty Ltd 495 Blackburn Road Mt Waverley Victoria 3149, AUSTRALIA
Manufacturer Single Registration Number (SRN):	AU-MF-000016740
European Representative:	CEpartner4U BV Esdoornlaan 13 3951 DB Maarn The Netherlands
European Representative Single Registration Number (SRN):	NL-AR-000000111
Product Name:	HistoCore PELORIS 3 and associated components listed in the attached Device Schedule A
Basic UDI-DI:	9349458002DB
Risk Class:	Class A – Rule 5 Annex VIII of Regulation (EU) 2017/746
Conformity Assessment Route:	Annex IV, in combination with Annex II and Annex III
Object of the declaration:	
Intended Use:	The HistoCore PELORIS 3 dual retort rapid tissue processor automates preparation of tissue samples for sectioning. This is achieved by transforming fixed specimens into wax infiltrated specimens by exposing them to a sequence of reagents in the tissue processor. Tissue samples subsequently undergo interpretation by a qualified healthcare professional to aid diagnosis.



The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:

Electromagnetic Compatibility
(2014/30/EU)

Restriction on the Use of Certain Hazardous
Substance in Electrical & Electronic
Equipment (2011/65/EU)

Waste electrical & Electronic Equipment
(2012/19/EU)

Regulation (EU) 2017/746 on In Vitro
Diagnostic Medical Devices

The following standards and technical documentation have been applied:

EN 61326-1:2013
(IEC 61326-1:2012 Edition 2.0)

Electrical equipment for measurement,
control and laboratory use- EMC
requirements. Part 1: General requirements.
Electrical equipment for measurement,
control and laboratory use-EMC
requirements- Part 2-6: Particular
requirements- In vitro diagnostic (IVD)
medical equipment.

EN 61326-2-6:2013
(IEC 61326-2-6:2012 Edition 2.0)

IEC 61010-1, Edition 2.0
UL/ IEC 61010-1, Edition 3.0

Safety requirements for electrical equipment
for measurement, control, and laboratory use
Part 1: General requirements

IEC 61010-2-010, Edition 2.0
IEC 61010-2-010, Edition 3.0

Safety requirements for electrical equipment
for measurement, control, and laboratory use
Part. 2-010, Particular requirements for
laboratory equipment for the heating of
materials

IEC 61010-2-081: 2001 (Edition 1) + A1:2003

Safety requirements for electrical equipment
for measurement, control, and laboratory use
Part 2-081: Particular requirements for
automatic and semi-automatic laboratory
equipment for analysis and other purposes

IEC 61010-2-101, Edition 1
IEC 61010-2-101, Edition 2

Safety requirements for electrical equipment
for measurement, control, and laboratory use
Part 2-101: Particular requirements for in vitro
diagnostic (IVD) medical equipment.

Signed for and on behalf of:

DocuSigned by:

Sandeep Chollangi



Signer Name: Sandeep Chollangi
Signing Reason: I approve this document
Signing Time: 21-Apr-2023 | 15:52:38 AEST

Sandeep Chollangi | ID: B02EB4A96186B

RA Manager
Leica Biosystems Melbourne Pty Ltd

SCHEDULE A

Component	Catalogue Number
HistoCore PELORIS 3 (220-240V)	45.0001
HistoCore PELORIS 3 (100-120V)	45.0005
Spaced Basket Kit	S45.4503
Basket Spaced	S45.4505
High Capacity Basket Kit	S45.4504
High Capacity Basket (with dividers)	S45.4506
HistoCore I-Scan Kit	S45.4507
HistoCore PELORIS 3 Barcode Scanner Kit	S45.4508

Revision No.	Date	Summary of Changes
A01	18 Feb 2022	Initial release and date of first compliance with Initial (EU) 2017/746
A02	06 June 2022	Update to add HistoCore I-Scan Kit and HistoCore PELORIS 3 Barcode Scanner kit
A03	21 Apr 2023	Update to include the intended use